

## **CHRISTIAN MEDICAL COLLEGE**

## DEPARTMENT OF CLINICAL BIOCHEMISTRY

## **CMC EXTERNAL QUALITY ASSURANCE SCHEME**





**Lab Name** Govt. PHC, Koratti Lab No 13046

Constituent Date of Result Entered: 20/08/2023 **Chemistry III** Group

PT item 01/09/2023 Lyophilized human serum based Date of Report Published:

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11	
SI.NO						CV	SD	Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	167 mg/dL	-0.95	0.42	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.78 mg/dL	0.17	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	108 mg/dL	-1.25	0.22	

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. PHC, Koratti

Koratti

**TAMILNADU** 

Panela Christudos

**Coordinator Contact Details:** Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss **CMC EQAS Coordinator Christian Medical College, Vellore** 

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components \*\*\*\*\*\* End of Report \*\*\*\*\*