



S.M. ENTERPRISES

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

27th
Years of Service

Office/Cal. Lab : Plot No. 12 "Ganesh Niwas", Sanjeevraah Nagar, Co-operative Society, Besides Tadibund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA / (040) 27840697, 48541008
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email msudheer 98@yahoo.com, msudheer 98@smentp.org, SME 003@YAHOO.IN, url www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/TIM/027/2023;Dt:25/05/2023
2. Requested by : (Name and address of the organization)	To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP Digital Timer
3. Instrument Name:	Aceteq / DT-1045
4. Instrument Make/Model:	SVMC/TIM/027
5. Instrument ID. No:	0 to 99 min
6. Range :	+/- 1% FSD
7. Accuracy :	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
8. Environmental Condition	Time V/S Time (Comparision Method)
9. Test parameter & condition :	Time Interval Meter
10. Standard Instruments used for Calibration:	C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023
11. Traceability Certificate Ref. No.	

TEST RESULTS

SI No	Standard Reading Of Time Hr:min:sec:ms	Observed Reading Of Time min:sec	Error Value(+/-) Hr:min:sec:ms
1	00:01:00:19	01:00	00:00:00:19
2	00:10:00:39	10:00	00:00:00:39
3	00:30:00:46	30:00	00:00:00:46
4	01:50:00:18	90:00	00:00:00:18

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
S. M. Enterprises



GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F
NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001-2015
TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS