

Centrifuge Calibration Verification Record Sheet

Location of the ICTC/PPTCT : ICTC, CHC, Bangarupalem, Chittoor District

Equipment ID: ICTC/BANGARUPALEM/CF/01

Range : up to 3500RPM

Procedure using centrifuge: In house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration: 19.09.2023

Due date of Calibration: 18.09.2024

Certified Tachometer Id: SRL/HIVTL/Tachometer

Certified Tachometer calibration date : 30.03.23

**Verification done by: G.Archana, Technical Officer,
Dept. of Microbiology, S.V. Medical College, Tirupati**


Sno	Standard reading (RPM)	UUC Reading (RPM)	Error (RPM)
1	499.7	500	0.3
2	999.5	1000	0.5
3	1998.7	2000	1.3
4	2999.5	3000	0.5
5	3497.1	3500	2.9

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge verification is satisfactory

Action taken: NA

Remarks: NA


Technical Officer
APSACS (SRL)
Department of Microbiology
S.V. Medical College/SVRRGGH
TIRUPATI - 517 507


Signature
Medical Officer
ICTC
Community Health Centre
BANGARUPALYAM - 517 416
Chittoor Dist. A.P. .