

		CALIBR	RATION	ERTI	FICATE			
CERTIFICATE NO:	SBS/CL/2	3/05131		The second secon	ECHANICAL	acceptable to the same of	Page No:1 of 1	
Issue Date		15-04-2023						
Receipt Date Calibration Date Calibration Due Customer Name & Address			SRF/23/00129-0003 & 13-04-2023					
			13-04-2023					
			12-04-20	12-04-2024				
GOVERNMENT URB		IFALTH CENT	TDE					
VADIVEESWARAM-	329002,NAGER	COIL.KANYAI	KLIMARI DISTRI	T				
			of Device Unde		n (DUC)			
		Dotails	of Device Office	Cambratic	on (DOC)			
escription : CENTRIFUGE			Make &	Make & Model : TAJAS ENTERPRISES				
Range :	3500	RPM	Sr. No		:	098.17		
Least Count	1	RPM	Identific	ation No	,			
DUC Condition : Satisfactory Location				97 : 1				
			ditions & Standa	ard Operati	ng Procedur	e Details		
Environmental Details Temperature		e: 25.8 °C Relative Hi		umidity	50% Rh			
Calibration Procedure No SBS/CP/ML/0		04 Calibration		done at ONSITE				
			Reference Stand	lards Detai	ls			
S.No Description		Make		ke/ SI No:		Certificate No		
Digital Tachometer L		LINE SEIKI / 175-0034V		TMS/22/63 01		10-11-2023		
CALIBRATION RES	ULTS				12 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1			

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	500	499.6	0.4	4.2
2	2000	1999.7	0.3	4.2
3	3000	2999.5	0.5	4.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By
(Calibration Engineer)
S.MURALI



Quality Manager (D.VETRI SELVI)

Chief Executive

Authorised Signatory

Galaxy~A33~5 $G_{
m INSHINE}$ biomedical solutions

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, Septemberat 2023 in 2:26 jam, Website: www.sunshinebiomedical.com.

1