



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/03727

Page No : 1 of 1

Customer Name & Address

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,  
VADIVESWARAM-629002,NAGERCOIL,KANYAKUMARI DISTRICT.

SRF No.	SRF/23/00129-0001
SRF Date	28-03-2023
Date of Receipt	28-03-2023
Date of Calibration	28-03-2023
Due Date for Calibration	27-03-2024
Issue Date	28-03-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	PRIMDA
Range	100-1000µl	Model	
Resolution	10µl	Material	PVC
Serial Number	111611023	Operating Range	
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.85	99.85	99.85	99.85	99.95	99.88	-0.12	0.05	0.47
		99.90	99.95	99.90	99.80	99.90				
2	500	499.90	499.95	499.90	499.90	499.95	499.88	-0.12	0.05	0.47
		499.80	499.80	499.85	499.85	499.90				
3	1000	999.80	999.80	999.80	999.85	999.90	999.87	-0.13	0.05	0.47
		999.90	999.90	999.90	999.90	999.95				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water containing Grade 3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)



Authorised by:

(Quality Manager/Chief Executive)  
D.VETRI SELVI

Galaxy A33 5G

21 September 2023 2:26 pm