

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. Block PHC, Karaiyur Lab No 7142

Constituent Date of Result Entered: 13/03/2023 **Chemistry III** Group

PT item Lyophilized human serum based Date of Report Published: 04/04/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI	
						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	98.43 mg/dL	1.02	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.35 mg/dL	0.17	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	112.6 mg/dL	0.29	0.28

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. Block PHC, Karaiyur The Block Medical Officer, Karaiyur, Pudukottai District Karaiyur TAMILNADU622402

> **Coordinator Contact Details:** Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss **CMC EQAS Coordinator Christian Medical College, Vellore**

Panela Christudoss

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report *****