

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Vellichandai

[Details About Robust Analysis](#)

Lab No 12268

[Detail About Monthly Summary](#)

Month July

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	105 mg/dL	-0.72	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	113 mg/dL	0.00	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)
