


CENTRIFUGE Calibration / Verification Record Sheet				
Location: ICTC-GGH-NANDYAL		Range : Up to 3500RPM		
Equipment ID: APSACS/ICTC/GGH/NDL/CEF/01		SL.NO :12612-2010		
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer				
Date of Calibration :	21.09.2023	Due date of Calibration:	20.09.2024	
Tachometer ID : KMC/MB/HIVTL/SRL/TM/01	Calibration date : 19.04.2023		Due date of Calibration : 18.04.2024	
Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC,KURNOOL				
S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)	
1.	504	500	-4.0	
2.	999.5	1000	0.5	
3.	1991.3	2000	9.7	
4	2998.7	3000	2.3	
5.	3499.1	3500	0.9	
Accept/ Reject: Accepted since average difference is within +/- 100RPM				
Final Interpretation: Centrifuge is in good working condition				
Action Taken: - NA				
Remark: - NA			 TECHNICAL OFFICER	