



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name **Govt. Block PHC, Kunichi** Lab No **3056**
 Constituent Group **Chemistry III** Date of Result Entered : **18/05/2023**
 PT item **Lyophilized human serum based** Date of Report Published : **05/06/2023**

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|----------------------------------------------|--------------------|--------|--------------|------|--------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 162.68 mg/dL | -1.94 | 0.44 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1854 | 1.76 | 7.48 | 0.13 | 1.83 mg/dL | 0.53 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1798 | 114.60 | 5.08 | 5.83 | 115.58 mg/dL | 0.17 | 0.27 |

| SDI Range | Interpretation |
|-----------------------|------------------------------------------|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. Block PHC, Kunichi
TIRUPATTUR TALUK, VELLORE DISTRICT, kunichi
Kunichi
TAMILNADU635601

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****