

CENTRIFUGE Calibration / Verification Record Sheet				
Location: ICTC- AH- AMALAPURAM, Dr. B.R. Ambedkar Konaseema District.			Range : Up to 3000 RPM	
Equipment ID: Remi ;			UID.NO :	
Calibration Procedure Reference: In-house verification of centrifuge using NABL calibrated Tachometer				
Date of Calibration: 22.9.2023			Due date of Calibration: 21.9.2024	
Certified centrifuge			Tachometer ID : EAPL ;DT – 2001B Calibration date : 7.6.2023 Calibration date due: 6.6.2024	
Verification Done by: S. Suneela Raja Kumari, Technical Officer, SRL, KMC, Kakinada				
S.NO	STANDARD READING (RPM)	UUC READING (RPM)	ERROR (RPM)	REMARK
1.	1700	1715	15	Acceptable
2.	2000	2018	18	Acceptable
3.	2300	2345	45	Acceptable
4	2700	2714	14	Acceptable
5.	3000	2934	-66	Acceptable
Accept/ Reject: Accepted since average difference is within +/- 100RPM				
Final Interpretation: Centrifuge is in good working condition				
Action Taken: - NA				
Remark: - NA				

State Reference Laboratory
Dept. of Microbiology
BMC, GGH Campus, KAKINADA

Suneela
20/9/2023
SRL-7-0