

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Killiyoor

Details About Robust Analysis

Lab No 4620

Detail About Monthly Summary

Month August

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	181 mg/dL	0.60	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22

SDI Range

Interpretation

Within -1.00 to +1.00

Excellent.

Within ±1.01 to ±2.00

Good.

Within ±2.01 to ±2.99

Accent with caution Warning Signal

Facility  
Location

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Killiyoor  
 Lab No 4620  
 Month July  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	107 mg/dL	-1.29	0.22

SDI Range

Interpretation

Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accent with caution Warning Signal

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Killiyoor

[Details About Robust Analysis](#)

Lab No 4620

[Detail About Monthly Summary](#)

Month June

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	292 mg/dL	0.36	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	109 mg/dL	-0.93	0.24
<b>SDI Range Interpretation</b>										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accent with caution Warning Signal						

**VIEW LAB MONTHLY SUMMARY**

Lab Name

Govt. Block PHC, Killiyoor

Lab No

4620

Month

May

Year

2023

Constituent Group

Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	182 mg/dL	0.12	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	2.1 mg/dL	2.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	120 mg/dL	0.93	0.27

**SDI Range**

**Interpretation**

Within -1.00 to +1.00

Excellent.

Within ±1.01 to ±2.00

Good.

Within ±2.01 to ±2.99

Accent with caution Warning Signal

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Killiyoor

Details About Robust Analysis

Lab No 4620

Detail About Monthly Summary

Month April

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	120 mg/dL	0.08	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Pictate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	2.3 mg/dL	-18.88	0.28

SDI Range

Interpretation

Within -1.00 to +1.00

Excellent.

Within ±1.01 to ±2.00

Good.

Within ±2.01 to ±2.99

Accent with caution Warning Signal

**VIEW LAB MONTHLY SUMMARY**

Lab Name Govt. Block PHC, Killiyoor  
 Lab No 4620  
 Month March  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 18/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	84 mg/dL	-1.10	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	107 mg/dL	-0.68	0.28

**SDI Range**

**Interpretation**

Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accent with caution Warning Signal