



# Metropolis Healthcare Ltd., EQAS

VITAL DIGITAL HEALTHCARE PVT LTD (719)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230104 and Sample No 03

Report Date : 19/05/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Horiba Yumizen H500

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.03	0.99	0.44	-2.77
@ Total Leucocytes (WBC) count	/c.mm	730.00	671.54	0.53	-0.59
@ Haemoglobin	g/dL	2.40	2.38	0.00	-3.18
@ MCV (Mean Corpuscular Volume)	fL	103.00	102.08	0.17	0.25
@ MCH (Mean Corpuscular Hb)	pg	23.50	24.09	-0.27	-0.58
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	22.80	23.62	-0.51	-0.50
@ Platelet Count	1000/ $\mu$ L	30.00	34.11	-0.68	-1.77
@ Hematocrit	%	10.60	10.14	0.60	-1.31
@ RDW CV%	%	20.00	20.84	-0.68	-0.14

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score  $\geq$ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070