

Notification No.	
Work Order No.	



SERVICE REPORT

ENGINEERS COPY

HIN-M0722-

CUSTOMER DETAILS	INSTRUMENT DETAILS		SERVICE STATUS
NAME: Vital Smart Lab	MODEL: Yumizen H500 EQUIP. SL. No.: 205Y0X404684 VERSION: 2.2.2.2b.		VISITING PURPOSE: <input type="checkbox"/> REPAIR <input type="checkbox"/> DATA ERROR <input type="checkbox"/> INSTALLATION <input checked="" type="checkbox"/> MAINTENANCE <input type="checkbox"/> UPGRADE <input type="checkbox"/> CUSTOMER TRAINING <input type="checkbox"/> COURTESY VISIT <input type="checkbox"/> DEMO <input type="checkbox"/> OTHERS
ADDRESS: Andheri (W)	NOTIFICATION No.		
PIN CODE 4000	CALL DETAILS	DATE	CUSTOMER STATUS <input type="checkbox"/> R.R <input checked="" type="checkbox"/> WARRANTY <input type="checkbox"/> AMC <input type="checkbox"/> CMC <input type="checkbox"/> DEMO <input type="checkbox"/> FREE SERVICE <input type="checkbox"/> CHARGEABLE CALL <input type="checkbox"/> OTHERS
CITY Mumbai	COMPLAINT RECD.	TIME	
STATE MH	START	14/04/23 15:00	WORKS CARRIED OUT AT <input checked="" type="checkbox"/> SITE <input type="checkbox"/> SERVICE CENTRE
TEL. NO.: 9952000205	COMPLETED	14/04/23 16:30	
CONTACT PERSON: Dr. Nilam N.	TRAVEL TIME (Eng. 1)	01 + 01 HOURS	
	TRAVEL TIME (Eng. 2)	HOURS	
	DAILY WORKLOAD	15+	

PROBLEM REPORTED: Maintenance

OBSERVATIONS:

ACTION TAKEN: System checked. done maintenance. adjust voltages. done concentrated cleaning startup pass. Sample run ok. Instrument working ok.

FOLLOWING PARTS HAVE BEEN REPLACED FOLLOWING PARTS NEED TO BE REPLACED PLEASE APPROVE

PART CODE.	DESCRIPTION (Replaced)	QTY.	COST	TAX	TOTAL
PART CODE.	DESCRIPTION (Trouble Shooting)	QTY.	TOTAL RS.		

*Replaced for Trouble Shooting
Need to be returned after Trouble Shooting.

FOLLOW-UP ACTION (Spare need to replaced, if any)		TO BE FILLED IN BY CUSTOMER	
		<input type="checkbox"/> FAULT RECTIFIED & INSTRUMENT IS WORKING SATISFACTORILY <input type="checkbox"/> FAULT IS NOT COMPLETED, ENGINEER NEED TO COME AGAIN <input type="checkbox"/> WE HEREBY APPROVE Rs. <input type="checkbox"/> COMMENTS (IF ANY)	
ENGINEER'S NAME-1: Yogesh G.	ENGINEER'S NAME-2:	VITAL SMART LABS Shop 1A, Ground Floor, Saptarishi Building, D.N.Nagar, Next to Joey's Pizza, Andheri (W), Mumbai - 53.	
SIGNATURE: <i>[Signature]</i>	SIGNATURE: <i>[Signature]</i>		
DATE & TIME:	DATE & TIME:		
		CUSTOMER SIGNATURE: NAME: Dr. Neelam Navagare	SEAL: