

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33ADSPA9153Q1ZA • DL No. TN-12-20B-00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Government Primary Health Centre, Valinokkam – 623528.	Certificate No:	AMS/LAB/289/3644
		Issuing Date:	11/05/2023
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Bio Chemistry Analyser	Id. No:	PHCV/LAB/BCA/01
Make:	Robonik	Department:	Laboratory
Model:	Prietest Touch	Calibrated On:	10/05/2023
Sl. No:	AT0380211 RBK	Calibration Due Suggested:	09/05/2024
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	CC23021536	31/01/2024
2	Manometer	Lutron/PM9100	PC/01/23/ATL/0512-04	31/01/2024
3	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
4	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	278	mA	500	Normal
2	Frequency	49.9	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	212.1	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	214.5	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	4.8	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	10.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	412.5	450	-17.5	µl
4	Peristaltic Pump pressure	-13.42	-15.0	-0.58	cmHZQ

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)