

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Vennavalkudi		Details About Robust Analysis							
Lab No	12544		Detail About Monthly Summary							
Month	April		Detail about SDI							
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 19/04/2023							
			Date of Report Published : 06/05/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124.3 mg/dL	0.67	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.92 mg/dL	-1.37	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	115.4 mg/dL	0.65	0.28
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out