



# MESO INSTRUMENTS

Calibration Laboratory  
Plot No. Plot A,B Sowmya Apartments,  
Kazura Garden, 1st Street, Neelangarai, Chennai - 600 115.



NABL ACCREDITED LABORATORY  
CC - 3089

## CALIBRATION CERTIFICATE

ULR NO: CC308920000001666F

|                    |   |               |             |
|--------------------|---|---------------|-------------|
| Customer & Address | M/s. The Medical Officer<br>Government Primary Health Center<br>Vennavalkudi, Pudukkottai Dist. | Cert.No       | MIT20/167-3 |
|                    |   | Date of Issue | 3-11-2022   |
|                    |   | Page No       | 1 of 1      |

### Detail of Device Under Calibration (DUC)

|              |                       |                    |            |
|--------------|-----------------------|--------------------|------------|
| Nomenclature | PIPETE                | Received Date      | 02-11-2022 |
| Make / Model | Microlux              | Cal. Date          | 02-11-2022 |
| SLNo         | -----                 | Due Date           | 01-11-2023 |
| ID. No       | ---                   | CALIBRATION STATUS | Functional |
| Range        | 100-1000 micro litter | Location           | LAB        |
| Resolution   | 0.1 micro litter      |                    |            |

### Standards used.

| SL No | Instrument  | Cert. No.            | Traceability                    | Valid up to |
|-------|---|----------------------|---------------------------------|-------------|
| 1     | Multifunction Calibrator<br>SL.No.0909A B40 1533A | ETDC(CN)/2019/52122  | ETDC,<br>Chennai                | 05-04-2023  |
| 2     | Weight box<br>SL.No.HTC17034567                   | RMTL/01/319107848-A1 | ROOTS<br>INDUSTRIES,<br>Chennai | 05-04-2023  |

Work instruction No: MI/WI/02

All calibration done in SI units and are traceable to National / International standards as required in ISO/IEC 17025:2017. This instrument has been calibrated at customer premises under the ambient conditions stated.

### Calibration Results

Temp: 25±4°C Humidity: 30 to 75%r.h

| SLNo | Standard Value<br>(micro litter)<br>(Avg.) | DUC Reading<br>(micro litter) | Error (°C) |
|------|--|-------------------------------|------------|
| 1    | 100  | 99.9                          | 0.1        |
| 2    | 200  | 199.9                         | 0.1        |
| 3    | 1000                                       | 999.9                         | 0.1        |

The expanded uncertainty to be associated with the results is calculated at a confidence level of approximately 95.45 %. (Coverage factor = 2)

### Remarks

1. This certificate shall not be reproduced, except in full, without written approval from the laboratory.
2. The recalibration interval should be determined based on the user's requirements.
3. The user should determine the suitability of the instrument for its intended use.
4. The results stated in this certificate relate only to the item calibrated.
5. The calibration points are taken as per customer requirement.

Calibrated by: S. Nagamani



E. Deenadayalan  
Authorised Signatory