8



BIO RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	DISTRICT HOSPITAL BHOPAL
Address:	Jay Prakash District Govt. Hospital Bhopal, 1250 link Road No.2 Tulsi Nagar Shivaji Nagar BHOPAL (M.P)
Province- Postal code	462001
Contact Name	Mr. Avneesh Kumar

Calibration Certificate

Date 01.09.2023 Instrument Model D-10Hb Analyzer Serial Number DM0G519494

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with (Device manufacturer and model	Device Serial Number	Device Calibration date (dd,mm,yyyy)	
Measuring cylinder hex base 10ML 1	1.20/2121	28.12.23	

Specified Range@1.5ml/Minute	Measured Value
1.425 – 1.575ml/minute	1.47

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with (Device manufacturer and model	Device Serial Number	Device Calibration date (dd,mm,yyyy)
FLUKE179	28400220	21.11.24
Specified Range@35°C	Measured Value	9
33.4°C – 36.4°C	35.2	şi

Detector log value

Reference: Service Manual: Section 7.2 or Field Bulletin D-10 - 081

Measured with	Specified Range	Measured Value	
Instrument Service Software	0.018 - 0.022AU	0.021AU	÷

The next calibration is due on	31.08.2024	(dd,mm,yyyy)

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory: Sourobh

Page1of1

Calibratics Cervilicate Property