

Certificate No : PHC/MP/001

CERTIFICATE OF CALIBRATION

Customer Details	Details of DUC	
PRIMARY HEALTH CENTRE, KILLIYOOR VILAVANCODE, KANNIYAKUMARI, TAMILNADU- 629157	Description	MICROPIPETTE
	Manufactured by	NIL
	Model	100-1000 μ l
	Serial No.	NIL
	Asset ID	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	27°C	Date of Receipt	26/05/2023
		Date of Calibration	28/05/2023
Humidity	56%	Calibration Due Date	27/11/2023
		Date of Issue	29/05/2023
		Cond. On Receipt	Satisfactory

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electronic Semi Micro Balance	AND, GR-202	14251938	27-SEP-2022	Equations

Performance Analysis

SI.No	Device Reading(μ l)	Mean Value(μ l)	Deviation
1.	1000	1000.1	0.1
2.	500	500.2	0.2
3.	100	99.9	0.1



Certificate No : PHC/MP/002

CERTIFICATE OF CALIBRATION

Customer Details	Details of DUC	
PRIMARY HEALTH CENTRE, KILLIYOOR VILAVANCODE, KANNIYAKUMARI, TAMILNADU- 629157	Description	MICROPIPETTE
	Manufactured by	NIL
	Model	10-100µl
	Serial No.	NIL
	Asset ID	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	27°C	Date of Receipt	26/05/2023
		Date of Calibration	28/05/2023
Humidity	56%	Calibration Due Date	27/11/2023
		Date of Issue	29/05/2023
		Cond. On Receipt	Satisfactory

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electronic Semi Micro Balance	AND, GR-202	14251938	27-SEP-2022	Equations

Performance Analysis

SI.No	Device Reading(µl)	Mean Value(µl)	Deviation
1.	100	100.1	0
2.	50	50.2	0.1
3.	10	9.9	0.1

Certificate No : PHC/MP/003

CERTIFICATE OF CALIBRATION

Customer Details	Details of DUC	
PRIMARY HEALTH CENTRE, KILLIYOOR VILAVANCODE, KANNIYAKUMARI, TAMILNADU- 629157	Description	MICROPIPETTE
	Manufactured by	NIL
	Model	5-50 μ l
	Serial No.	111610688
	Asset ID	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	27 $^{\circ}$ C	Date of Receipt	26/05/2023
		Date of Calibration	28/05/2023
Humidity	56%	Calibration Due Date	27/11/2023
		Date of Issue	29/05/2023
		Cond. On Receipt	Satisfactory

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electronic Semi Micro Balance	AND, GR-202	14251938	27-SEP-2022	Equations

Performance Analysis

Sl.No	Device Reading(μ l)	Mean Value(μ l)	Deviation
1.	10	10.1	0
2.	5	5.2	0.1
3.	1	0.9	0.1