

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Authoor Lab No 12976

Constituent Group

Chemistry III

Date of Result Entered :

Panela Christudoss

20/07/2023

PT item Lyophilized human serum based

Date of Report Published : 07/08/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	U
SI.NO						CV	SD	Value	וטפ	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	115 mg/dL	0.93	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	55 mg/dL	0.63	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.5 mg/dL	-0.94	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS :
Govt. PHC, Authoor
GOVT PRIMARY HEALTH CENTRE
Authoor

TAMILNADU628151

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******

1 of 1 9/25/2023, 7:07 PM