



CALIBRATION /TESTING REPORT

CUSTOMER DETAILS : UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE, REDDIARPATTI, TIRUNELVELL, TAMILNADU-627007		CERTIFICATE NO	TICS/2023/RDPI/010
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	LABOTECH
		MODEL	BD-153
CALIBRATION DATE	09/05/2023	SERIAL NO	76706
CALIBRATION DUE DATE	08/05/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875

TEST REPORT Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 2	2234	± 5	PASS
		Speed 3	2805	± 5	PASS


*The above mentioned instrument has been tested using standard manufacturer recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By :

[Handwritten Signature]

Authorized Sign With Seal



[Handwritten Signature]

*Electrical safety test result attached