

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/11634	MEDICAL DEVICES	Page No:1 of 1
Issue Date	19-09-2023	
SRF No & Date	SRF/23/00391-0003 & 16-09-2023	
Receipt Date	16-09-2023	
Calibration Date	16-09-2023	
Calibration Due	15-09-2024	

Customer Name & Address

GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE,
PARTHIBANUR-623608.

Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : ATCD3561220RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 97	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 175	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 235	Uncertainty in % (±) 7.2

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised Signatory

CSW
Technical Manager
C.SHANMUGARAJ

Chief Executive