



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---------------------------------------------------------------------------------------------------------------|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 230623A/N-117 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 24 / 06 / 2023 | Ref. Std. | : CG-11 | |
| CALIBRATION DUE. | : 23 / 06 / 2024 | C.P. No. | : ACS/CP/07 | |
| SRF No. & DATE | : 230623A/N Dt. 23.06.2023 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC34272300000579F Dt. 26.06.2023 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House No.2(Dazy House) Sarvodya Nagar, Udaypur Khas, Bareilly-243001 | | | |
| ITEM CALIBRATED | : Digital Temperature Indicator | | | |
| | Make : RoHS | Range | : (-50 to 110) °C | |
| | L. Count : | ID No. | : RCL/SR/D.Temp-04 | |
| | Location : Sample Receiving Dept. | S. No. | : N.A | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C} R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2022/0846 | 10/11/2023 |
| RESULTS | : ET Discipline | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.1 | 2 | 0.1 | | |
| 4.2 | 4 | 0.2 | | |
| 6.2 | 6 | 0.2 | | |
| 8.3 | 8 | 0.3 | | |
| (average of readings) | | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

