


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

|                                 |   |                        |               |                    |
|---------------------------------|---|------------------------|---------------|--------------------|
| <b>CERTIFICATE NO.</b>          | : 230623A/N-103   | <b>Page</b>            | : 1 of 1      |                    |
| <b>CALIBRATED ON.</b>           | : 23 / 06 / 2023  | <b>Ref. Std.</b>       | : ISO:8655:6  |                    |
| <b>CALIBRATION DUE.</b>         | : 22 / 06 / 2024  | <b>C.P. No.</b>        | : ACS/CP/08   |                    |
| <b>SRF No. &amp; DATE</b>       | : 230623A/N Dt. 23.06.2023  | <b>Phy. Inspection</b> | : O.K.        |                    |
| <b>ULR No. &amp; ISSUE DATE</b> | : CC342723000000565F Dt. 26.06.2023   | <b>Cal. At</b>         | : Site / Lab  |                    |
| <b>CALIBRATED FOR</b>           | : M/s Redcliffe Lifetech Pvt. Ltd.<br>House No.2(Dazy House) Sarvodya Nagar,<br>Udaypur Khas, Bareilly-243001 |                        |               |                    |
| <b>ITEM CALIBRATED</b>          | : Micro Pipette   |                        |               |                    |
|                                 | <b>Make</b>   | : Thermo Scientific    | <b>Range</b>  | : (100 to 1000) µl |
|                                 | <b>L. Count</b>   | :                      | <b>ID No.</b> | : RCL/HEM/PIP-03   |
|                                 | <b>Location</b>   | : Haematology          | <b>S. No.</b> | : TW05908          |
| <b>ENVIRONMENT CONDITION:</b>   | Temp. { 25°C ± 3°C } R.H. {40% to 60%}  |                        |               |                    |

| STANDARD INSTRUMENTS USED : |                       |              |                 |            |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME                 | ID No. / S. No.       | TRACEABILITY | CERTIFICATE No. | DUE DATE   |
| Weighing Balance            | ACS/DB/02/ D449927064 | BlueBox      | BBT/398/JAN/23  | 05/01/2024 |

| RESULTS : Mechanical Discipline |                 |                    |                      |            |
|---------------------------------|-----------------|--------------------|----------------------|------------|
| Nominal Capacity (µl)           | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
|                                 |                 | 100                | 100.35               | (-) 0.35   |
| 1000                            |                 | 500                | 501.66               | (-) 1.66   |
|                                 |                 | 1000               | 1003.29              | (-) 3.29   |

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

AS  
Prepared by

....End of Calibration Certificate....

