



	ELECTRICAL S	SAFETY TESTING/ CALIBRATION	
Report No : HBMS/CAL/23-24		Calibration Date: 31.08.2023	
		Calibration Valid: 03.03.2024	
1.1CUST	OMER DETAILS:		
Name and Address of the organization		M/S, GOVT, PRIMARY HEALTH CENTRE,	
		KABILARMALAI, NAMAKKAL.	
1.2. DESC	CRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME		MICROSCOPE	
i	Manufactured by	LABOMED	
i ii	Manufactured by Model	LABOMED BINACULAR MICROSCOPE	

2.CONDITION OF THE ITEM WHEN RECEIVED		
No visible damages and in working order		

3.TEST DONE	Electrical safety/ Performance Testing
-------------	--

4. STATUS						
Manufactures Specification	User specification	Within specification	Out of specification	Calibration	Electrical Safety Test	Performance Analysis
1	-	1	-	-	V	

V

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

0

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com





5.TRACEABILI	TY DETAILS OF INSTR	RUMENTS USED 1	FOR TESTING		
S.no	Name of the instrument	Make	Model	Serial no	Cal Due
1.	Electrical safety analyser	Fluke	Esa612	3005050	07-OCT-2023

S.NO	PARAMETER	OBSERVED VALUE	ACCEPTANCE LIMITS AS PER THE STD	REMARKS
1.	Protective earth resistance	0.129Ω	<0.3 Ω	Ok
2.	Earth leakage	134.21μΑ	<210 µA NC	Ok
3.	Patient leakage current	NA.	<100 μA BF	Ok
4.	Equipment current	0.3A	As per the manufacture spec	Ok
5.	Mains voltage (P-E)	230.5V	-	Ok
6.	Mains voltage (P-N)	231.2V	-	Ok
7.	Mains voltage (N-E)	1.0V	<5V	Ok
8.	Enclosure leakage current	50 μA	<100 μA (normal)	Ok
9.	Insulation resistance	OL	>100 M Ω	Ok

S.NO	7.Remarks	
1.	This report is applicable to the sample tested only	
2.	The instrument used for testing are valid calibration and are traceable to national standards	
3.	Parameter of the DUT were verified and found to be within the specified limits	

Verified by:

V

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com

GSTIN: 33AAJFH1833M1ZB

MD Reg No: TN/CBE/MD42/00016