



CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 31.08.2023
		Calibration Valid: 03.03.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVT, PRIMARY HEALTH CENTRE, KABILARMALAI, NAMAKKAL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MICROLUX
II	Model	-----
III	Value	100µl
IV	Serial No	GPHC/KBI/MP/002

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	100µl
1	100.05
2	100.06
3	100.01
4	99.97
5	100.01
6	100.02

For Heart Beat Medical Systems,

(Authorized signatory)

