



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. Block PHC, Siruvalur  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 7370  
Date of Result Entered : 20/06/2023  
Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	258 mg/dL	-1.93	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	110 mg/dL	-0.74	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. Block PHC, Siruvalur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.  
Siruvalur  
TAMILNADU638054

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

மருத்துவ அலுவலர்,  
அரசு ஆரம்ப சுகாதார நிலையம்,  
சிறுவலூர்.

Pamela Christudoss  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential