



REPORT ON  
ELECTRICAL SAFETY TESTING/ PERFORMANCE ANALYSIS/ CALIBRATION

Report No : TR/SPHC/021/23-24  
Page 1 of 2

Calibration Date : 18/02/2023  
Calibration Due : 17/02/2024

1.1 CUSTOMER DETAILS

Name and address of the organisation	MEDICAL OFFICER GOVT PRIMARY HEALTH CENTER, Perundurai main road, Siruvalur, Erode Tamil Nadu - 638054
Reference and Date	Letter dated 18-02-2023
Date of receipt of item	16/02/2023'

1.2 DESCRIPTION OF DEVICE UNDER TEST (DUT)

NOMENCLATURE	BIO CHEMISTRY ANALYZER
Manufactured by	ROBONIK
Model	PRIETEST TOUCH
Serial No.	ATCD1680519RBK
Biomedical Product ID	SPHC/LAB/BCA/01
Supply	220V-240V AC
Device Type	Type CF
Device Classification	Class II equipment
Location	LABORATORY

1.3 CONDITION OF THE ITEM WHEN RECEIVED

No visible damage and in working order

1.4 ENVIRONMENTAL CONDITION OF MEASUREMENTS

A.Temperature	28.6°C
B.Relative Humidity	45-75%
C.Ambient Barometric Pressure	756mmHg

1.5 Applicable Specification IEC Specification IEC 60601-1, IEC 60601 -2-27

1.6 Test Done Electrical Safety and Performance Testing

Tested by :

Approved by :

Balaji Krishnan K  
(Biomedical Engineer)

Priya M  
(Quality Manager)



FL BIOMEDICAL  
189, Vasantham Paradise,  
Chithode, Erode-638102.  
Cell: 7092848995