

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/11012

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
SENUR-632006,VELLORE DISTRICT.

SRF No.	SRF/23/00360-0016
SRF Date	06-09-2023
Date of Receipt	06-09-2023
Date of Calibration	06-09-2023
Due Date for Calibration	05-09-2024
Issue Date	07-09-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	ERBA BIOHIT
Range	10µl FIXED	Model	NA
Resolution	0.2µl	Material	PVC
Serial Number	7067133	Operating Range	10µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.8	Humidity (%RH)	56	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		1	2	3	4	5				
1	10	9.92	9.40	9.96	9.97	9.93	9.86	-0.14	0.17	0.47
		9.87	9.88	9.89	9.91	9.90				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

C. Sivabalan
(Calibration Engineer)
C.SIVABALAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.