



## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/00222	MEDICAL DEVICES	Page No:1 of 1
Issue Date	13-01-2023	
SRF No & Date	SRF/23/00009-0002 & 12-01-2023	
Receipt Date	12-01-2023	
Calibration Date	12-01-2023	
Calibration Due	11-01-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
RAMAPURAM-603201,MADURANTHAKAM TALUK,CHENGALPATTU DISTRICT.

**Details of Device Under Calibration (DUC)**

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH	
Range :	Sr. No : AT1391206MRK	
Resolution :	Identification No :	
DUC Condition : SATISFACTORY	Location : LABORATORY	

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature: 25.3° C	Relative Humidity	54% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	TSC/22-23/7400-3	10-08-2023

**ELECTRICAL SAFETY**


**RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 96	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 131	Uncertainty in % (±) 5.9
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 118	Uncertainty in % (±) 11.4

**REMARKS**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
M.BALAJI

Authorised Signatory

  
Quality Manager  
(D.VETRI SELVI)

Chief Executive

