

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Ramapuram	Details About Robust Analysis
Lab No	12226	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/08/2023


Date of Report Published : 01/09/2023

Sl.No	Analyte	Method Name	Principle	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
							CV	SD			
1	GLUCOSE III	GOD-POD III		Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	180 mg/dL	0.49	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III		Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.82 mg/dL	0.50	0.01
3	CHOLESTROL III	CHOD-PAP III		Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	108 mg/dL	-1.25	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
**ASSISTANT SURGEON**  
**DVT. PRIMARY HEALTH CENTRE**  
**Ramapuram - 603201**  
**Chengalpattu District**