



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name	PATHOLOGY, DISTRICT HOSPITAL NEEMUCH	Lab No	7809
Constituent Group	HbA1c	Date of Result Entered :	18/07/2023
PT item	Lyophilized human whole blood based	Date of Report Published :	07/08/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Biorad (D10 / D100 / Variant Turbo)	711	6.66	3.98	0.26	6.6 %	-0.23	0.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
PATHOLOGY, DISTRICT HOSPITAL NEEMUCH
DISTRICT HOSPITAL NEEMUCH MP
NEEMUCH
MADHYA PRADESH458441

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****