



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	DISTRICT HOSPITAL NEEMUCH
Address:	District Govt. Hospital, Mahu Nasirabad Road Neemuch (M.P)
Province- Postal code	458441
Contact Name	Mr. Mujahid Hussain

Calibration Certificate

Date	24.03.2023	Instrument Model	D-10Hb Analyzer	Instrument Serial Number	DM0H527158
------	------------	------------------	-----------------	--------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with (Device manufacturer and model)	Device Serial Number	Device Calibration date (dd,mm,yyyy)
Measuring cylinder hex base 10ML	2310925	03.10.24
Specified Range@1.5ml/Minute	Measured Value	
1.425 – 1.575ml/minute	1.53	

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with (Device manufacturer and model)	Device Serial Number	Device Calibration date (dd,mm,yyyy)
FLUKE179	26030158	24.06.24
Specified Range@35°C	Measured Value	
33.4°C – 36.4°C	34.9	

Detector log value

Reference: Service Manual: Section 7.2 or Field Bulletin D-10 - 081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022AU	0.021AU

The next calibration is due on	23.03.2024	(dd,mm,yyyy)
--------------------------------	------------	--------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

Sourabh

Calibration Certificate

