



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. Block PHC, Muttom
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 7392
 Date of Result Entered: 19/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	177.49 mg/dL	0.21	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	110.22 mg/dL	-0.75	0.81
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Muttom
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, KANYAKUMARI DISTRICT.
 Muttom
 TAMILNADU629202

Coordinator Contact Details:
 Email: cilingc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

[Signature]
GOVT. PRIMARY HEALTH CENTRE, MUTTOM

Homogeneity and Stability of the sample is passed.
 Data In CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Muttom

Details About Robust Analysis

Lab No 7392

Detail About Monthly Summary

Month July

Detail about SDI

Year 2023

Constituent Chemistry III
Group

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method - Principle Name	Analyzer Host	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	107.5 mg/dL	-0.30	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.84 mg/dL	0.03	0.02
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	114.68 mg/dL	0.39	0.80
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)


[View Lab Monthly Summary](#)

[Verify Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)


MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
MUTTOM

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Muttom
 Lab No 7392
 Month June
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	288.79 mg/dL	0.15	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	116.77 mg/dL	0.47	0.76

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

R. N. S.
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 MUTTOM