

VIEW LAB MONTHLY

Govt. Block PHC, Muttom

Lab Name 7302

Lab No June

Month 2023

Year

Constituent Chemistry III

Group

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	288.79 mg/dL	0.15	0.6
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.0
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	149	114.61	4.05	4.64	116.77 mg/dL	0.47	0.7

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

*R. N. S.*

MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
MUTTOM