



Lab Name Govt. Block PHC, Muttom

Lab No

T392

Constituent Group Chemistry III

Date of Result Entered :

20/04/2023

PT item Lyophilized human serum based

Date of Report Published :

06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	CV	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	116.52 mg/dL	-0.41	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.03 mg/dL	-0.74	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79	108.43 mg/dL	-0.57	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Muttom
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, KANYAKUMARI DISTRICT.
Muttom
TAMILNADU629202

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
GOVT. MUTTOM

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components