

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Muttom
 Lab No: 7392
 Month: March
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 18/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	83.48 mg/dL	-1.17	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.31 mg/dL	-0.17	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78	115.69 mg/dL	0.87	0.79

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Early Summary	Change Password	Home	Sign-out

SDI : Standard Deviation Index

Chief
19/4/2023

R. Altranpau
4/4/2023
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENT.
MUTTOM