


Certificate No : GPHCM/002



**BIOMEDICAL
TECHNIQUES**
Calibration experts

CALIBRATION CERTIFICATE

Customer Details	Details of DUC	
GOVT PRIMARY HEALTH CENTRE, MUTTOM, KANYAKUMARI, TAMILNADU- 629202	Description	CENTRIFUGE
	Manufactured by	MICROSIL
	Model	MK-10
	Serial No.	226/385
	Identification No.	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	28°C	Date of Receipt	07/03/2023
Humidity	55 %	Date of Calibration	07/03/2023
		Date of Issue	10/03/2023
		Due Date of calibration	06/03/2024
		Calibrated At	Site

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electrical safety analyser	Fluke / ESA 615	2244039	25-SEP-2023	Accuver Callibration Laboratory CO LLC, Dubai
Digital Tacho Meter	Amprobe	BMT TACHO 031	30/09/2023	Equation

Performance Analysis

Parameter	Set Value (RPM)	Measured Value(RPM)	Tolerance	Error Observed
RPM	1000	1001	±5 %	1
	2000	2002	±5 %	2