

Certificate No : GPHCM/001



CALIBRATION CERTIFICATE

Customer Details	Details of DUC	
GOVT PRIMARY HEALTH CENTRE, MUTTOM, KANYAKUMARI, TAMILNADU- 629202	Description	MICROSCOPE
	Manufactured by	LABOMED
	Model	VISION 2000
	Serial No.	160892674
	Identification No.	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	28°C	Date of Receipt	07/03/2023
Humidity	55 %	Date of Calibration	07/03/2023
		Date of Issue	10/03/2023
		Due Date of calibration	06/03/2024
		Calibrated At	Site

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electrical safety analyser	Fluke / ESA 615	2244039	25-SEP-2023	Accuver calibration laboratory CO LLC, Dubai

Electrical Safety Test

Parameter	Measured Value	Tolerance	Result
Ground wire Resistance	0.156 Ω	<0.3 Ω	Pass
insulation Resistance	OL	> 2 M Ω	Pass
Chassis Leakage Current *NC	0.6 μ A	<100 μ A	Pass
Chassis Leakage Current *SFC	0.4 μ A	<500 μ A	Pass
Patient Leakage Current - B	0.8 μ A	<100 μ A	Pass