

## CENTRIFUGE CALIBRATION VERIFICATION RECORD SHEET

<b>Location of the ICTC/PPTCT : ICTC, CHC- B.KOTHAKOTA, CHITTOOR DISTRICT</b>			
<b>Equipment ID: ICTC/ B.KOTHAKOTA /CF/01</b>		<b>Range : up to 3500RPM</b>	
<b>Procedure using centrifuge: In house verification of centrifuge using certified standard calibrated by Tachometer</b>			
<b>Date of Calibration: 13-09-2023</b>			
<b>Due date of Calibration: 12-09-2024</b>			
<b>Certified Tachometer Id: SRL/HIVTL/Tachometer</b>			
<b>Certified Tachometer calibration date : 30-03-2023</b>			
<b>Verification done by: G.ARCHANA, Technical Officer, Dept. of Microbiology, S.V.Medical College, Tirupati</b>			
S.No	Standard reading (RPM)	UUC Reading (RPM)	Error (RPM)
1	499.0	500	1.0
2	998.4	1000	1.6
3	1999.2	2000	0.8
4	2998.5	3000	1.5
5	3498.3	3500	1.7
<b>Accept/ Reject: Accepted since average difference is within +- 100RPM</b>			
<b>Final Interpretation: Centrifuge verification is satisfactory</b>			
<b>Action taken: NA</b>			
<b>Remarks: NA</b>			

*Archana*  
**Signature**  
**Technical Officer**  
**APSACS (SRL)**  
**Department of Microbiology**  
**S.V. Medical College/SVRRGGH**  
**TIRUPATI - 517 507**