

## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/03261

Page No: 1 of 1

**Customer Name & Address**

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,  
NO.1.BHAJANAI KOVIL STREET,SEMBAKKAM,TAMBARAM

SRF No.	SRF/23/00111-0014
SRF Date	20-03-2023
Date of Receipt	20-03-2023
Date of Calibration	20-03-2023
Due Date for Calibration	19-03-2024
Issue Date	20-03-2023

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	MICROLUX
Range	100-1000µl	Model	
Resolution	10µl	Material	PVC
Serial Number		Operating Range	
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

**Environmental Condition**

**Calibration Method Used**

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

**Z Factor: 1.00319**

**Result of Calibration in µl**

SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	100.51	100.52	100.53	100.52	100.51	100.50	0.50	0.02	0.47
		100.49	100.48	100.49	100.49	100.48				
2	500	500.40	500.42	500.43	500.44	500.46	500.43	0.43	0.02	0.47
		500.45	500.44	500.43	500.42	500.41				
3	1000	1001.24	1001.26	1001.23	1001.22	1001.22	1001.22	1.22	0.03	0.47
		1001.18	1001.17	1001.20	1001.24	1001.26				

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming to Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)

R.YAZINIYAN



Authorised by:

(Quality Manager/Chief Executive)

D.VETRI SELVI