Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company





GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer	M/S. Community Health Centre,	Certificate No:	AMS/LAB/289/3060
	Dharapuram Road, Vellakoil-638111.	Issuing Date:	20/02/2023
Address:		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	PHCV/LAB/VMP/002
Make:	Finnpipette F'3	Department:	Laboratory
Model:	Thermo	Calibrated On:	18/02/2023
SI. No:	PW10778	Calibration Due	
Type:	Variable	Suggested:	17/02/2024
Range &	10 – 100μ1	DUC Condition	
Least Count:	• 50	on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test ResultsVolume µl:

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(µl)	Trial 1	Trial 2	(µl)
1	10.0	10.00	10.00	00.00
2	20.0	20.00	20.00	00.00
2	50.0	50.00	50.00	+00.00
3	100.0	105.00	105.00	+05.00

Allowable Tolerance: User Defined

Note

- 1. Calibration Points were selected as per customer request.
- 2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
- 3. Results Reported are valid at the time of and underthe stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

amshealthcaremetre phail.com

Approved by:

(Technical Manager)

☐ 94434 74610, 93455 56148

