


CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-104

Lab Name: Govt. Block PHC, Nallur
 Chemistry III
 PT Item: Lyophilized Neuman serum based
 Lab No: 7442
 Date of Result Entered: 21/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Mean Value	SD	CV
1	GLUCOSE III	GOO-P00 III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	8.13	9.80	176.1 mg/dl	0.06	0.42
2	CREATININE III	Jaffe Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.75 mg/dl	0.09	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1784	113.80	4.07	4.04	114.8 mg/dl	0.17	0.22
SCN Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.01 to +2.00				Good.						
Within +2.01 to +2.98				Accept with caution. Warning Signal.						
Beyond +3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Nallur
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NAMAKKAL DISTRICT,
 Nallur
 TAMILNADU637203

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EOAS Coordinator
 Christian Medical College, Vellore

A. K. S.
 மருத்துவ கட்டுப்பாட்டு
 திட்டமிடல்

Homogeneity and Stability of the sample is passed.
 Data in CMC EOAS reports is confidential