



CALIBRATION CERTIFICATE

CUSTOMER ADDRESS : M/s. THE BLOCK MEDICAL OFFICER(TPM364/10),
 GOVT PRIMARY HEALTH CENTRE,
 KOTHANALLUR,
 MEKKAMANDAPAM POST,
 KANYAKUMARI 629166
 Phone No.:

UNIQUE LAB REPORT NUMBER : ULR-CC231723000003530F	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT : MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : TPMCC230859241	DATE OF ISSUE : 10-08-2023
RECEIPT NO : 12033	TEMPERATURE : 23.4 °C
RECEIPT DATE : 07-08-2023	HUMIDITY : 52 %
JOB CARD NO : TPM364082357664	AIR PRESSURE : 960 hpa
JOB CARD DATE : 07-08-2023	MEDIUM OF WEIGHING : GRADE III WATER
BRAND : FINNPIPETTE - F3	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 10 -100 µl	DENSITY CORRECTION FACTOR [Z] : 1.0035 µl/mg
SERIAL NO : RW01329	CALIBRATION METHOD : ISO 8655:2002 PART - 6
TIPS USED : 200 µl	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE (1 µl - 10000 µl)	

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT : SHIMADZU
 SL.NO : D450020184
 RESOLUTION : 0.00001g upto 82g & above is 0.0001g
 CALIBRATED BY : THE PIPETEMANN (MASS), COIMBATORE
 CERTIFICATE NUMBER : TPMCCM23020418
 CALIBRATION VALIDITY : 03-02-2024
 TRACEABILITY : NATIONAL STANDARDS