

## VIEW LAB MONTHLY SUMMARY

|                   |                    |                               |  |
|-------------------|--------------------|-------------------------------|--|
| Lab Name          | Govt. UGPHC,Muthur | Details About Robust Analysis |  |
| Lab No            | 7620               | Detail About Monthly Summary  |  |
| Month             | August             | Detail about SDI              |  |
| Year              | 2023               |                               |  |
| Constituent Group | Chemistry III      |                               |  |

Click on the analyte to view Graphical Data

Date of Result Entered : 18/08/2023


Date of Report Published : 01/09/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value  | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|-------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SB   |             |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1841               | 175.58 | 5.13         | 9.00 | 168.9 mg/dL | -0.74 | 0.42 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1852               | 1.76   | 6.77         | 0.12 | 1.78 mg/dL  | 0.17  | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1764               | 113.80 | 4.07         | 4.64 | 115.4 mg/dL | 0.35  | 0.22 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

## Links

|                |                 |                    |                          |
|----------------|-----------------|--------------------|--------------------------|
| Modify Address | Modify Methods  | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home               | Sign-out                 |

  
**மருத்துவ அலுவலர்,**  
**அரசு மேம்படுத்தப்பட்ட**  
**ஆரம்ப சுகாதார நிலையம்,**  
**முத்தூர், திருப்பூர் மாவட்டம்-638 105.**