



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Name: Govt. PHC, Alamarathupatty

Lab No

12134

Constituent Group: Chemistry III

Date of Result Entered :

06/07/2023

Method: Lyophilized human serum based

Date of Report Published :

07/08/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
					CV	SD		
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	121 mg/dL	1.92
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	4.87 mg/dL	-2.74
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	101 mg/dL	-2.58

SDI Range	Interpretation
-1.00 to +1.00	Excellent.
±1.01 to ±2.00	Good.
±2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :

PHC, Alamarathupatty
MEDICAL OFFICER, ATHOOR
Alamarathupatty
TAMIL NADU

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****