

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name **Govt. PHC, Kaverinagar**Lab No **9227**[Back to Search Page](#)Month **July**Year **2023**[Print](#)Constituent Group **Chemistry III**Date of Result Entered : **21/07/2023**Date of Report Published : **07/08/2023**

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10			1.70
			Transasia / Erba	24	109.83	6.83	7.50			3.06
			Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	112 mg/dL	0.44	0.28
			Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95			1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34			0.12
			Transasia / Erba	18	5.79	3.00	0.17			0.08
			Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.5 mg/dL	-0.94	0.02
			Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27			0.07
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04			1.39
			Transasia / Erba	19	112.64	4.38	4.93			2.26
			Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	108 mg/dL	-1.07	0.22
			Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88			0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

