



**ROOTS INDUSTRIES INDIA LIMITED**  
**ROOTS METROLOGY & TESTING LABORATORY**  
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India  
 Phone :9003432356



NABL/ILAC/024

NABL ACCREDITATION  
 CERTIFICATE NO CC-2752

## CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

<b>Name and address of the customer</b> The Medical Officer Government Primery Health Centre Kumbeswaram south uphc –thanjavur (dist)	<b>Certificate no</b> : Rmtl/01/319107849/04 <b>Date of calibration</b> : 27/09/2023 <b>Suggested calibration due</b> : 26/09/2024 <b>Customer po .reference.&amp; date:</b>
--	---

### 1.DESRIPTION OF DUC

<b>Name Of The Duc:</b> CENTRIFUGE <b>Make</b> :BDI INSTRUMENT <b>Model</b> : BDI-152 <b>Sl.No</b> : NA <b>Range/Size</b> : 1-4000RPM/230v	<b>Receiving Condition Of Duc</b> :Good <b>Ref. Document</b> :As Per Manufactures Specification <b>Environmental Condition</b> :25.3 Deg & 54.3% Rh <b>Uncertainty Of Measurement</b> :As Per Report <b>Location</b> :lab
--	---

**THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.**

### 2.DETAILED OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2024
Digital Tachometer	Mextec	Rml-251	Gsg/19/206/Tm/01	1-6-2024
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2024
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2024

**CALIBRATION FIELD:** Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 230v	230	231	1	0.055

**CURRENT CALIBRATION**Ac Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

**Resistance calibration**

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

**CALIBRATION FIELD:** RPM Calibration

Sl.no	Range	Master readings RPM	Calibrated values RPM	Error observed RPM	Uncertainty RPM
1	1RPM to 4000RPM	1600RPM	1599RPM	1RPM	0.055

### 4.remarks

4.1 uncertainty estimated for a confidence level of 95% at the k factor=2

4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration

4.3 DUC –device under calibration

Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore, cell:9585840192
--------------------------	----------------------	---------------------	---