



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 27.09.2023
		Calibration Valid: 26.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, PERIYAPOTHU,COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BDI-153
iii	Serial No	76040

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	2	1.12
2	1200	1201	2	1.23
3	1500	1502	4	1.56
4	2000	1999	1	1.175

Acceptable deviation is ± 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016



GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662



No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com