



**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24		Calibration Date: 27.09.2023
		Calibration Valid: 26.03.2024
<b>1.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE, PERIYAPOTHU, COIMBATORE</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MERILETTE
II	Model	5µl - 50µl
III	Serial No	22212239
IV	Location	LABORATORY

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.05	20.08	50.97
2	05.06	20.07	20.02
3	05.01	20.05	20.96
4	05.97	20.99	20.02
5	05.01	20.08	20.09
6	05.02	20.95	20.04

**For Heart Beat Medical Systems,**

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com