

Chenna 600 03

		CALIBR	ATIO	V CERTIFICA	ATE		
CERTIFICATE NO: SBS/CL/23/09989				MEDICAL DEVICES		Page No:1 of 1	
				11-08-2023			
Issue Date SRF No & Date				SRF/23/00305-0004 & 10-08-2023			
Receipt Date				10-08-2023			
Calibration Date				10-08-2023			
Calibration Due				09-08-2024			
	mer Name & Address		100-00				
-	RNMENT URBAN PRIMA	ARY HEALTH CENTRE	5 1 10 10 10 10 10 10 10 10 10 10 10 10 1			The state of the state of the state of	
	IESWARAM-623701,PAR						
LIVIAIN	12017/10/10/-023701,1 7/10	Details	of Device l	Jnder Calibration (DUC)	the state of the		
Description : SEMI AUTO ANALYSER Range : MULTI			Make Sr. No	& Model :	ALPHA & ALPHA CHEM-100 BCAA17052570		
Resolution : MULTI			Identif	ication No :	NA		
DUC Condition : SATISFACTORY				Location		BORATORY	
		Environmental		& Calibration Procedure			
Environmental Details Temperature: 25.7° C				Relative Humidity 48% R			
Calibration Procedure No SBS/CP/MD/20			Calibration done at ONSITE		y		
154			Reference S	Standards Details			
S.No Description Make/ SI No			l No:				
1 Electrical Safety Analyser FLUKE & 22			k 2244202	14202 TSC/22-23/5114-2 23-06-2024			
RESI	JLTS		Elect	rical Safety 🧀 💹			
S.no	Specification	Measured values in MΩ		Allowable limit in MΩ Unce		linty in % (±)	
1	Insulation Resistance	92		>20ΜΩ		13.92	
		Measured values in		Allowable limit in μA	Uncertainty in % (±)		
2	Earth Leakage	140	* 1 pv.	<5000µAfor B,BF,CF	8.9		
	Measured values in		in µA	Allowable limit in µA	Uncertainty in % (±)		
3	Enclosure Leakage	217		<500µAfor B,BF,CF	8.8		
2.The 3.The 4.The 5.The	ARKS s Calibration certificate shall user should determine the recalibration interval should results stated in this certific indicated uncertainties are uipment used for Calibration Calibrated By	suitability of the instrume d be determined on the us cate relate only to the item expanded uncertainty es a were calibrated & tracea	nt for its inte ser requirent calibrated. timated for a	ended use. nent. a confidence level of appro nal & International Standa	oximately 95% for a cov		
	(Calibration Engineer) K.SATHYAMOORTHY	W Change	<u>v</u>	Technical M	Manager NMUGARAJ)	Chief Executive	